

## Seal Checker

Packaging						
	9	Sterintech™	Seal	Checker	<b>REF</b> 909.0	001.0250
	[]i	Place the Seal Checker into a pouch and seal it by the heat sealer. The area sealed should be darker and	Location: Sealer: Date:		M 11-20 LOT 0111 ▲	
	ш	uniform SP Medikal San Ltd	Result: PA	SS FAIL	T	
	****	SP Medikal San Ltd	Hesuit: PA	SS THAIL	11	

## Release your Heat Sealer for the day...

The ISO-11607 standard is specifying a routine, daily check of your heat sealers at the beginning of the day. A way to perform the test required and to document which sealer, which day, which person did perform the test and to easily show results of your heat sealer capacity of properly heat seal your pouch the Sterintech Seal Checker is a very convenient aid.

When place inside a pouch and sealed the sealed area will be darker and can as such be easily read whether the seal is consistent, regular and completely without any gaps etc.

Seal Checker REF: 909.001.0250 Unit: 250 sheets

## Characteristics:

- ◆ Dimensions: 175 x 75 mm.
- ◆ Suitable for for all rotary heat sealers.
- ◆ Can be fully recycled.
- ◆ Conforms to ISO 11607-2

Other types and dimensions available upon request.







Tel: +90-212-613 80 54

Fax : +90-212-613 80 55 (EUR) Account No.: 9089601 Email: info@spmedikal.com (US\$) Account No.: 9089602 Web: www.spmedikal.com VAT: Bayrampaşa - 781 046 3595

Bank: Garanti Bank - Istanbul - TURKEY BIC : TGBATRIS Branch No.: 437 IBAN: TR36 0006 2000 4370 0009 0896 01 IBAN: TR09 0006 2000 4370 0009 0896 02 Mersis No: 0781046359500019

